




## TOF-CIMS

### Chemical Ionization Time-of-Flight Mass Spectrometer

*Real-time chemical analysis of trace gases, aerosols, or atmospheric ions.*



#### APPLICATIONS

- Online identification and quantification of trace gas- or particle-phase compounds
- Laboratory, field, or mobile platform based experiments
- Air quality and climate change research

#### ADVANTAGES

- Quantitative response with broad dynamic range
- Sub pptv gas-phase limits of detection
- Molecular identification and elemental speciation
- Interchangeable ionization sources for selective detection of different chemical classes
- High-speed data acquisition
- Low power, field portable assembly

#### INTERCHANGEABLE INLETS AND IONIZATION SOURCES

The ToF-CIMS is used with three different chemical ionization sources, which can be easily interchanged by the user.

- **Flow Tube Ion Molecule Reaction Chamber (IMR):** Reduced, tuneable pressure (30 to 500 mbar) reaction ionization chamber for use with acetate, iodide, and water cluster reagent ions. Compatible with Po or X-ray ionizer. The standard configuration uses a gas-phase inlet. The IMR can optionally be used with the **FIGAERO Inlet**, which enables real-time chemical analysis of gas and aerosols.
- **Atmospheric Pressure Drift Tube Reaction Chamber:** For use with nitrate reagent ion. Measurement of highly oxidized gas-phase organic compounds. Not compatible with FIGAERO.
- **Extractive Electrospray Ionization Source (EESI):** Ionization source for online molecular analysis of aerosols.

## TOF Mass Analyzer

The ToF-CIMS is available in two models that differ in size and mass resolving power.

	Mass Resolving Power (M/ $\Delta$ M)	Dimensions
HToF-CIMS	4000 - 6000	59 x 42 x 83 cm
LToF-CIMS	7000 - 9000	61 x 48 x 151 cm

## Performance and Specifications

- **Gas-Phase LOD:** For example, 4 pptv formic acid (1 s, Bertram, 2011), 0.4 pptv malonic acid (15 s, Lee, 2014)
- **Aerosol LOD, FIGAERO:** For example, 4 ng/m<sup>3</sup> formic acid and 2 ng/m<sup>3</sup> C<sub>9</sub>pinene acid (Lopez-Hilfiker, 2014)
- **Aerosol LOD, EESI:** 1 ng/m<sup>3</sup> dipentaerythritol and raffinose (Lopez-Hilfiker and Slowik, 2019)
- **Mass Range:** 0 to 1000 Th
- **Save Rate:** Up to 200 complete mass spectra/second
- **Detection polarity:** Bipolar TOF mass analyzer
- **Power:**
  - ToF-CIMS: <2 kW peak, <1.3 kW steady state
  - FIGAERO: < 0.5 kW
- **Sample Flow Rate:**
  - IMR: 2 lpm
  - FIGAERO: 2 lpm gas, 5 - 15 lpm aerosol
  - Atm Pressure Drift Tube: 10 lpm
  - EESI: 1 lpm

## Software

Tofware is an Igor-based post processing software, with workflows for quantification and high-resolution peak fitting and identification.

## REFERENCES:

### Ion Molecule Reaction Chamber (IMR)

A field-deployable chemical ionization time-of-flight mass spectrometer Atmos. Meas. Tech., 4, 1471-1479. doi:10.5194/amt-4-1471-2011, 2011.

Particulate Organic Matter Detection Using a Micro-Orifice Volatilization Impactor Coupled to a Chemical Ionization Mass Spectrometer (MOVI-CIMS), Aerosol Sci. Technol., 44(1), 61-74, doi:10.1080/02786820903380233, 2010.

An iodide-adduct high-resolution time-of-flight chemical-ionization mass spectrometer: application to atmospheric inorganic and organic compounds, Environ. Sci. Technol., 48(11), 6309-6317, doi:10.1021/es500362a, 2014.

### Atmospheric Pressure Nitrate Ionization Source

Measurement of the gas phase concentration of H<sub>2</sub>SO<sub>4</sub> and methane sulfonic acid and estimates of H<sub>2</sub>SO<sub>4</sub> production and loss in the atmosphere Geophys. Res., 98(D5), 9001-9010, doi:10.1029/93JD00031., 1993.

Atmospheric sulphuric acid and neutral cluster measurements using CI-APi-TOF Atmos. Chem. Phys., 12, 4117-4125, doi:10.5194/acp-12-4117-2012, 2012.

A large source of low-volatility secondary organic aerosol Nature 506, 476-479 doi:10.1038/nature13032

### Real-time Aerosol Measurements (FIGAERO)

A novel method for online analysis of gas and particle composition: description and evaluation of a Filter Inlet for Gases and AEROsols (FIGAERO), Atmos. Meas. Tech., 7(4), 983-1001, doi:10.5194/amt-7-983-2014, 2014.

An Extractive Electrospray Ionization Time-of-Flight Mass Spectrometer (EESI-TOF) for online measurement of atmospheric aerosol particles, Atmos. Meas. Tech. Discuss., <https://doi.org/10.5194/amt-2019-45>, in review, 2019.